

# Manual Supplement

|                |                        |                   |      |
|----------------|------------------------|-------------------|------|
| Manual Title:  | 1550C/1555 Calibration | Supplement Issue: | 2    |
| Print Date:    | September 2010         | Issue Date:       | 1/15 |
| Revision/Date: |                        | Page Count:       | 1    |


---

---

This supplement contains information necessary to ensure the accuracy of the above manual.

## Change #1, 65440, 65598

On page 4, add the following to the **Symbols** table:

|   |  |
|---|--|
|  | Conforms to relevant South Korean EMC Standards.   |
| CAT II  | Measurement Category II is applicable to test and measuring circuits connected directly to utilization points (socket outlets and similar points) of the low-voltage MAINS installation. |
| CAT III   | Measurement Category III is applicable to test and measuring circuits connected to the distribution part of the building's low-voltage MAINS installation.                               |
| CAT IV  | Measurement Category IV is applicable to test and measuring circuits connected at the source of the building's low-voltage MAINS installation.   |

On page 19, under **General Specifications**, remove the **Enclosure Setting, Electromagnetic Compatibility, Certifications, Safety Compliance** and **Pollution Degree 2** and add:

- IP Rating ..... IEC 60529: IP40
- Safety ..... IEC 61010-1: 600 V CAT IV / 1000 V CAT III, Pollution Degree 2
- Electromagnetic Environment ..... IEC 61326-1: Portable
- Electromagnetic Compatibility ..... Applies to use in Korea only. Class A Equipment (Industrial Broadcasting & Communication Equipment)  
 This product meets requirements for industrial (Class A) electromagnetic wave equipment and the seller or user should take notice of it. This equipment is intended for use in business environments and is not to be used in homes.

## Change #2, 110

On page 12, replace Table 6 with:

**Table 1. Voltage Measurement Test**

| Step | Voltage Source Output | UUT Display     | UUT Tone | UUT Display Limits |         |
|------|-----------------------|-----------------|----------|--------------------|---------|
|      |                       |                 |          | Minimum            | Maximum |
| 1    | -38 Vdc               | Flashing Hazard | Beeps    | 30 V               | 46 V    |
| 2    | 240 Vac, 60 Hz        | Flashing Hazard | Beeps    | 202 V              | 278 V   |