

JCB_F_12.02 2012-02



CERTIFICATE

No. Z1 13 07 14544 098

Holder of Certificate:

Fluke Corporation

FLUKE.

6920 Seaway Boulevard Everett, WA 98203-9090

USA

Production Facility(ies):

56588

Certification Mark:



Product: Multimeters

(Insulation Multimeter)

Model(s): 1577, 1587, 1587C

Parameters: Rated Input Voltage: 6VDC

(4 x 1.5V AA batteries IEC LR6)

Protection Class:

1000V CAT III and 600V CAT IV Measurement Category:

IP40 Degree of Protection: Temperature Ambient: 55°C

Tested according to: EN 61010-1:2010

EN 61010-2-030:2010

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

PI1305384108-000 Test report no.:

Date, 2013-07-22

Page 1 of 1



