

Manual Supplement

Manual Title: 1507/1503 Users
Part Number: 2416013
Print Date: June 2005
Revision/Date:

Supplement Issue: **6**
Issue Date: 10/14
Page Count: 3

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1

On page 4, under **Battery Saver (*Sleep Mode*)**, replace the last sentence in the first paragraph with the following:

The Tester comes out of Sleep mode when the rotary switch is turned to off and back on.

On page 24, under Model 1503, replace the Display Range, Output Voltage 500V:

From: 0.1 to 20.00 M Ω

To: 0.01 to 20.00 M Ω

Change #2


On page 5, under Table 2, add the following:

Warning

To avoid shock, injury, or damage to the Tester, use only accessories with 1000V CAT III or 600V CAT IV rating.

Change #3

On page 3, add the following to the **Symbols** table:

	Conforms to relevant South Korean EMC Standards.
©	Conforms to CAN/CSA-C22.2 No. 61010-1, second edition, including Amendment 1.
CAT II	Measurement Category II is applicable to test and measuring circuits directly to utilization points (socket outlets and similar points) low-voltage MAINS installation.
CAT III	Measurement Category III is applicable to test and measuring circuits connected to the distribution part of the building's low-voltage MAINS installation.
CAT IV	Measurement Category IV is applicable to test and measuring circuits connected at the source of the building's low-voltage MAINS installation.

Change #4, 61173

On page 21, under **General Specifications**, replace **Altitude** with:

Altitude..... Operating: 2000 m CAT IV 600 V, Non Operating (Storage): 12,000 m

Change #5, 64919, 65176, 367

On page 20, under **General Specifications**, remove **Shock, Electromagnetic Compatibility, Safety, Certifications**, and add:

- Safety** IEC 61010-1: 600 V CAT IV, Pollution Degree 2
- Electromagnetic Environment** IEC 61326-1: Portable
- Electromagnetic Compatibility** Applies to use in Korea only. Class A Equipment (Industrial Broadcasting & Communication Equipment)

This product meets requirements for industrial (Class A) electromagnetic wave equipment and the seller or user should take notice of it. This equipment is intended for use in business environments and is not to be used in homes.